

Weightron Bilanciai Ltd,
Titan Works
Bridge Way off Broombank Road,
Chesterfield Trading Estate,
Chesterfield S41 9QJ, UK
Tel: 01246 260062
Fax: 01246 260844
www.weightron.com



Weightron Bilanciai Ltd,
Unit 4b, Block 4
Westpark, Chelston
Wellington TA21 9AD, UK
Tel: 01823 662355
Fax: 01823 662017
www.weightron.com



CALIBRATION CERTIFICATE

This Calibration Certificate is issued in accordance with the UKWF Code of Practice for the calibration of Non Automatic Weighing Machines. All weights used are traceable to NPL

Customer

Ellgia Recycling
Winterton Road
Scunthorpe
LOCATION DN15 0DH

Machine Details

Cust ref	
Make	Bilanciai
Model	DD700
Ser No	T213102741
Capacity	50/000 / kg
e/d	/ 20/kg
Working cap	

Test Weights

Set No.	Cert No.	Date Cal
WB1-22	9445	Apr 22

Cert Interval	12 mth
Next Cert	MAR 2024

Tolerance +/-	kg
0- 10000	20
10000-40000	40
40000-50000	60

Comments	2 3
.....	
.....	
.....	1 4
.....	Ⓚ

AS FOUND LINEARITY & HYSTERESIS kg

Load	Loading	Dev	Unloading	Dev
Zero	0	0	0	0
2000	2000	0	2000	0
4000	4000	0	4000	0
8000	8000	0	8000	0
18000	18000	0	18000	0
28080	28080	0	28080	0
38080	38080	0	38080	0
43720	43740	+20	43740	+20

DEFINITIVE LINEARITY & HYSTERESIS

Load	Loading	Dev	Unloading	Dev

Repeatability kg

Load	Ind	Ind	Ind
2000	2000	2000	2000
20080	20080	20080	20080
43720	43740	43740	43740

Repeatability

Load	Ind	Ind	Ind

Eccentric Load 9640 kg

Reading	Dev	Reading	Dev
1 9660	+20	2 9640	0
3 9640	0	4 9640	0
5		6	
7		8	

Eccentric Load

Reading	Dev	Reading	Dev
1		2	
3		4	
5		6	
7		8	

Signature A.W. Lewis Name A.W. Lewis Date 31 MAR 2023

C1.3.88.1N