

**CALIBRATION CERTIFICATE**

Certificate No	88127
----------------	-------

Contact Name:	Darrel McKinney
Customer Name:	Ellgia Recycling Ltd
Site Address:	Unit 7 Lancaster Way Ely Cambridgeshire CB6 3NW
Email:	Darrel.mckinney@ellgia.co.uk; stephen.baggaley@ellgia.co.uk


All Test Weights / Test Rigs used are traceable to National Standards.

LINEARITY		
Applied Weight	Reading As Found	Reading As Left (Definitive)
0 kg	N/A	0 kg
2000 kg	New Indicator	2000 kg
10000 kg		10000 kg
20000 kg		20000 kg
30000 kg		30000 kg
44080 kg		44100 kg

EQUIPMENT DETAILS	
Make / Model:	Dini Argeo 3590 ETP
Serial Number:	74035139
Customer ID No.	N/A
Location:	Site
Capacity:	50t
Division Size:	20 kg

CALIBRATION DETAILS	
Accuracy:	+/- 0.1%
Test Weights Certificate Number(s)	TR10905, TK0057
Calibration Date:	24 July 2020
Calibration Due:	July 2021

ECCENTRIC LOADING		
Corner 1	N/A	10000 kg
Corner 2	New Indicator	10000 kg
Corner 3		10000 kg
Corner 4		10000 kg

CERTIFICATE ISSUED BY	
Name:	Ellis Blyth
Signature:	
Date:	28 July 2020

QF1 – Revision 3