

CALIBRATION CERTIFICATE

Certificate No	090970
----------------	--------

Contact Name:	Stephen Baggaley
Customer Name:	Ellgia Skips
Site Address:	Padnal Sidings Ely Road Prickwillow Ely CB7 4UJ
Email:	stephen.baggaley@ellgia.co.uk

All Test Weights / Test Rigs used are traceable to National Standards.

LINEARITY		
Applied Weight	Reading As Found	Reading As Left (Definitive)
0 kg	0 kg	0 kg
6000 kg	6000 kg	6000 kg
11000 kg	11000 kg	11000 kg
16000 kg	16000 kg	16000 kg
20000 kg	20000 kg	20000 kg
30000 kg	30000 kg	30000 kg
44100 kg	44080 kg	44080 kg

EQUIPMENT DETAILS	
Make / Model:	Precia Molen X112-B
Serial Number:	01F912425
Customer ID No.	N/A
Location:	Site
Capacity:	50,000 kg
Division Size:	20 kg

CALIBRATION DETAILS	
Accuracy:	+/- 0.1%
Test Weights Certificate Number(s)	TJ0005, TR10905
Calibration Date:	01 July 2020
Calibration Due:	July 2021

ECCENTRIC LOADING		
Corner 1	7000 kg	7000 kg
Corner 2	7000 kg	7000 kg
Corner 3	7000 kg	7000 kg
Corner 4	7000 kg	7000 kg

CERTIFICATE ISSUED BY	
Name:	Ellis Blyth
Signature:	
Date:	07 July 2020

QF1 – Revision 3